



+91.22.22057323
+91.84879 47881

Product Specification

WATER TESTING KIT

(For the estimation of Residual Chlorine)

CAT. NO. : KIT026

GRADE : LR

CAS NO. : _

FORMULA : _

M.W. : _

ISO 9001:2015 CERTIFIED



- Take 5ml of water sample in the test bottle(No-1)
- Add 5 drops of Residual Chlorine reagent-1 (RC-1) and mix well.
- The colour is compared with the Residual Chlorine colour chart and record the residual chlorine value.

Lab Reagents

DATE: 01.07.2025

Remarks: The above-mentioned batch material is compliance with the standard specifications of High Purity Laboratory Chemicals Pvt. Ltd.



Customercare@hplcindia.com



hplcindia.com