ISO 9001:2015 CERTIFIED HPLC Lab Reagents
320, Gold Mohur Bldg., 174, Shamaldas Gandhi Marg, Mumbai - 400 002 (India)
Plot No. 1902, Fansa Road, G.I.D.C., Sarigam – 396 155, Dist. Valsad, Gujarat (India)
+91.22.22057323
+91.84879 47881

Product Specification

UNIVERSAL INDICATOR PAPER PH RANGE 1-14

CAT. NO. : PH0097
GRADE : LR
CAS NO. : _
FORMULA : _
M.W. : _

Sensitivity	Passes test

ISO 9001:2015 CERTIFIED HPLC Lab Reagents

pH-Indicator Paper – Usage Note

- Slight discoloration of dry, unused papers may occur due to their high sensitivity to pH and ambient exposure. This does **not** affect their performance in pH determination.
- In weakly buffered or unbuffered solutions, **indicator error** may arise. To minimize this:
  1. Adhere the strip to the **inner wall of a test tube**.
  2. Fill the tube up to the top edge of the paper with the test solution.
  3. Wait **30 seconds to 1 minute**.
  4. Compare the paper's color with the reference scale **through the glass**.

DATE: 01.07.2025

Remarks: The above-mentioned batch material is compliance with the standard specifications of High Purity Laboratory Chemicals Pvt. Ltd.